

# Is your state of affairs in MDR *state of the art*?

**Michelle Lott, RAC**

Principal & Founder, leanRAQA

# MEDICAL DEVICE QUALITY IS ALL WE DO, AND WE'RE ALWAYS AHEAD OF THE GAME.

1.5M

years industry  
experience

522k


podcast listeners

182k+

look to us for the  
latest in quality

#1

blog and podcast  
in the industry



“Best eQMS I have ever  
used...”

This is the easiest eQMS I have used in  
the 20 years I have been in the Medical  
Device Industry. ***It is simple, intuitive  
and easy to use...*** We are successfully  
implementing a Quality Culture.

- Director of Regulatory Affairs  
& Quality Assurance

“Modern QMS Software and Outstanding Customer Service.”

★★★★★

“Demystifying QMS and Regulatory Requirements”

★★★★★

“Makes your QMS Simple and Effective”

★★★★★


# About leanRAQA



Expert in RAQA for ~ 20 years  
Advisor on FDA GMP Advisory Committee  
Master Grief Counselor

**Michelle Lott**

Principal and Founder



**Morty**

Chief Entertainment Officer



**Chopper**

Chief Fun Officer



**Lucy**


Chief Therapy Officer

**State of the Art (SOTA)**  
**vs.**  
**Standard of Care (SOC)**

---

*More than semantics?*


# At the heart of the argument



# Lifecycle of State of the Art



# SOTA according to...



**What say you, Mister ISO 14971:2019?**



# ISO 14971:2019 says... [Definition]



1<sup>st</sup> time in history

**State of the Art**: Developed stage of technical capability [...] as regards products, processes and services, based on the relevant consolidated findings of science, technology and experience.

- Currently and generally accepted as good practice in technology and medicine
- Does not imply the most technologically advanced solution

Mentioned 15 times

# ISO 14971:2019 says... [Pre-market]



The manufacturer takes into account the generally acknowledged state of the art, in order to determine the suitability of a medical device to be placed on the market for its *intended use*.

Information related to the generally acknowledged *state of the art* can include new or revised standards, published validated data specific to the application of the medical device under consideration, the availability of alternative medical devices and/or therapies, and other information **(10.2)**


# ISO 14971:2019 says... [Post-market]




The manufacturer should also take into account considerations of the generally acknowledged state of the art, including new or revised standards. **A 2.10**

The conditions under which follow-up actions need to be considered, are extended with changes in the state of the art that can be relevant to safety, such as alternative medical devices and/or therapies becoming available on the market, as well as changes in risk perception or risk acceptability. **A 2.10**

# Lifecycle of State of the Art



# SOC according to...



# Wherefore art thou, Standard of Care?



**What say you, Mister MDR?**

# MDR says...



Medical Devices Regulation  
EU MDR

**70 new definitions**


Does **NOT** define:

- State of the Art
- Standard of Care

**Mentioned SOTA 11 times**



# What MDR does say – Ch. VIII Cooperation w/ expert labs



## What MDR does say – Annex 1 General Safety and Performance

- ▶ Devices shall achieve the performance intended that... they are suitable for their intended purpose. They shall be safe and effective ... provided that any risks which may be associated with their use constitute acceptable risks when weighed against the benefits ... taking into account the generally acknowledged state of the art.
- ▶ Risk control measures ... for the design and manufacture of the devices shall conform to safety principles, taking account of the generally acknowledged state of the art....


## What MDR does say – Annex XIV Clinical Eval. and PMCF


- ▶ To plan, continuously conduct and document a clinical evaluation, manufacturers shall establish and update a clinical evaluation plan which shall include at least:
  - An indicative list and specification of parameters to be used to determine, based on the state of the art in medicine, the acceptability of the benefit-risk ratio;


# What MDR does say – Annex XV Clinical Investigations

- ▶ The clinical investigation plan (CIP) shall set out ...
  - The current state of the art in clinical care in the relevant field of application and the proposed benefits of the new device.
  - Description of the relevance of the clinical investigation in the context of the state of the art of clinical practice.
- ▶ The Sponsor shall prepare a clinical investigation report which includes the following:
  - Discussion and overall conclusions covering safety and performance results, assessment of risks and clinical benefits, discussion of clinical relevance in accordance with clinical state of the art....

# Interchangeable concepts?







# **What say you, Mister MEDDEV 2.7.1/4?**

**Clinical Evaluation: A guide for manufacturers and notified bodies  
under Directives 93/42/EEC and 90/385/EEC**



# MEDDEV 2.7.1/4 says...



Medical Devices Guidance  
MEDDEV

Does **NOT** define:

- State of the Art
- Standard of Care

**Mentioned SOTA 39 times**

# General principles of clinical evaluation (1 of 4)

- ▶ ... Determination of the benefit/risk ... acceptability of that profile based on current knowledge / the state of the art in the medical fields concerned.
- ▶ Verify...a high level of protection of health and safety and acceptable according to current knowledge / the state of the art

## General principles of clinical evaluation (2 of 4)

### ► Definition of scope

- The current knowledge / state of the art in the corresponding medical field, such as applicable standards and guidance documents, information relating to the medical condition managed with the device and its natural course, benchmark devices, other devices and medical alternatives available to the target population.

- Data on the safety and performance of other devices and alternative therapies, including benchmark devices and equivalent devices, should be used to define the state of the art or identify hazards due to substances and technologies.

## General principles of clinical evaluation (3 of 4)

- ▶ Expert documents ... are important for assessment of current knowledge / the state of the art, including clinical practice guidelines and consensus statements.
- ▶ Importance of literature review to risk management process. The literature review will provide data on current interventions for the intended patient population (state of the art) in order to give input to the assessments of acceptable benefit/risk profiles, what is currently considered as providing a high level of protection of health and safety and what are considered acceptable side-effects.

## General principles of clinical evaluation (4 of 4)

- Evaluate if the clinical data on benefits and risks are acceptable for all medical conditions and target populations covered by the intended purpose when compared with the current state of the art in the corresponding medical field and whether limitations need to be considered for some populations and/or medical conditions.

# **MDCG 2020-6**

**Clinical evidence needed for medical devices previously CE marked under Directives 93/42/EEC or 90/385/EEC. A guide for manufacturers and notified bodies**

# MDCG 2020-6 says...



## 8 new definitions

**DOES** define:

- State of the Art

Does **NOT** define:

- Standard of Care

**Mentioned SOTA 20 times**

**Mentioned SOC 4 times**

## MDCG 2020-6

### ► IMDRF/GRRP WG/N47 provides the following definition:

- Developed stage of current technical capability and/or accepted clinical practice in regard to products, processes and patient management, based on the relevant consolidated findings of science, technology and experience.
- Note: The state-of-the-art embodies what is currently and generally accepted as good practice in technology and medicine. **The state-of-the-art does not necessarily imply the most technologically advanced solution.**




## MDCG 2020-6

- ▶ The common features of the devices which are “well- established technologies” (WET) are that they all have:
  - relatively simple, common and stable designs with little evolution;
  - their generic device group has well-known safety and has not been associated with safety issues in the past;
  - well-known clinical performance characteristics and their generic device group are standard of care devices where there is little evolution in indications and the state of the art;
  - a long history on the market.

## MDCG 2020-6

- ... for low risk standard of care devices where there is little evolution in the state of the art, it may be possible to demonstrate conformity with the relevant GSPRs with a more limited clinical data set.



## MDCG 2020-6

- ▶ ... in practise the data collected may not meet MDR criteria, if the devices were considered **standard of care** and were not associated with safety concerns. Stable, **(WET)** that perform as intended and are **not associated with safety concerns**, and where there **has been no innovation**, are less likely to be the subject of research, and therefore literature data may be limited or non-existent. ... **may be necessary ...to undertake PMCF** ... prior to certification under the MDR, **even if they are (WET) and have been on the market for several decades**, to enable an evaluation of their safety and clinical performance in relation to **an evolving state of the art**.

## MDCG 2020-6

- In exceptional cases, particularly for low risk standard of care devices where there is little evolution in the state of the art, and the device is identified as belonging to the group of 'well-established technologies' ... a lower level of clinical evidence may be justified to be sufficient for the confirmation of conformity with relevant GSPRs.

## MDCG 2020-6

### ► Clinical State of the Art v Technical State of the Art

#### **Clinical SOTA:**


Relating to  
evaluation of clinical  
data

#### **Technical SOTA:**

Relating to current  
technological  
capability

**How do I establish New State of the Art and  
Standard of Care?**

# Step 1





# What to expect from the CECP or expert panels?



Created by EC to support scientific assessment, deliver opinions and provide ad hoc advice



Who are they? CVs and Dols:  
[https://ec.europa.eu/health/md\\_expert\\_panels/experts/expert\\_panels\\_en](https://ec.europa.eu/health/md_expert_panels/experts/expert_panels_en)

Appointed for 3 years, with possibility of renewal

Checks if there's a need for a scientific opinion

Screening panel

Orthopaedics, traumatology, rehabilitation, rheumatology

Circulatory system

Neurology

Respiratory system, anaesthesiology, intensive care

Endocrinology and diabetes

General and plastic surgery

Obstetrics and gynaecology, including reproductive medicine

Gastroenterology & hepatology

Nephrology & urology

Ophthalmology

IVD

First opinion published on June 16, 2021  
[https://ec.europa.eu/health/md\\_expertpanels/list-opinions-cecp\\_en](https://ec.europa.eu/health/md_expertpanels/list-opinions-cecp_en)

12 panels


# First opinion – More transparency?

## ► NB0483, Ivory Graft Ltd, 2021-000201\_NB0483\_opinion:


- Following a request from the notified body the opinion has been temporarily removed.

MDC Medical Device  
Certification GMBH

“The new device offers an interesting addition to the available portfolio of devices for the claimed indications. However, presented **clinical data** for one indication (#4) **are presently insufficient** and should be extended to include at least the healing phase for the implant (additional 4 months) and the results can then also be used for a positive clinical assessment for indication #3 and #5. For the other indications data from **clinical studies are missing** and therefore the evidence for these indications is insufficient. **Literature survey is flawed by the fact that the new device is similar but not equivalent to market products.** The **PMCF plan needs to be extended and specified** even to cover in detail the other claimed indications. If relevant data are available, indications can be accordingly extended.”



## Step 2



# What is a Common Specification (CS)?




Set of technical and/or clinical requirements – other than a standard – that provide a means of complying with the legal obligations.

Applicable to a device, process, or system when no harmonized standards exist, when relevant harmonized standards are not sufficient, or when there is a need to address public health concerns.

**MM 202X?**


**Are we there yet?**

# Lifecycle starts over



**Back to the heart of the argument**

# At the heart of the argument






# SOTA and your Technical Documentation


- MDD v MDR Technical Documentation Requirements – significant additional requirements (particularly for clinical evaluation process)




# At the heart of the argument



# SOTA and your Risk Analysis




# At the heart of the argument



# SOTA and your Post Market Surveillance


- Requirement of PMS activities to show your device is still based on SOTA or whether it has changed.



# New Post-Market Surveillance Requirements


[MDCG 2020-5](#)

**Guidance on Clinical  
Evaluation - Equivalence**




[MDCG 2019-9 Rev.1](#)

**Summary of Safety and Clinical  
Performance**



[MDCG 2020-6](#)

**Guidance on Sufficient Clinical  
Evidence for Legacy Devices**



[MDCG 2020-7](#)

**Guidance on PMCF plan  
template**




[MDCG 2020-8](#)

**Guidance on PMCF evaluation  
report template**



# At the heart of the argument




# Clinical documentation supersedes TD





# SOTA and your Clinical Evaluation: CER is king



# New Clinical Requirements

[MDCG 2021-28](#)

**Substantial modification  
of clinical investigation  
under Medical Device  
Regulation**



[MDCG 2021-20](#)

**Instructions for generating CIV-ID  
for MDR Clinical Investigations**

[MDCG 2021-8](#)

**Clinical investigation  
application/notification documents**

[MDCG 2020-7](#)

**Regulation (EU)  
2017/745 – Questions &  
Answers regarding  
clinical investigation**

[MDCG 2020-13](#)

**Clinical evaluation  
assessment report template**

# New Technologies Guidance



## [MDCG Infographic:](#)

Is Your Software a Medical Device?



## [MDCG 2021-5](#)

Guidance on standardization of medical devices



## [MDCG 2020-1](#)

Guidance on clinical evaluation (MDR) / Performance evaluation (IVDR) of medical device software



## [MDCG 2019-11](#)

Qualification and classification of software



## [MDCG 2019-16 Rev.1](#)

Guidance on cybersecurity for medical devices

# Example: SOTA Digital Imaging System

- Note: “state of the art” refers only to products that are developed and approved for sale in the marketplace.
- Difference between a new state-of-the-art digital imaging system that is undergoing trials and one that already has CE Marking.
  - For EU medical device regulators, the latter is considered state of the art, but the former is not until it has CE Marking.

X-Ray Machine

Considered state-of-the-art




X-Ray Machine with Remote Monitoring Features (under review, not approved yet)



**What does this mean for you?**


# Are we having a heart attack yet?



# Are we having a heart attack yet?

- ▶ Perform an in-depth assessment of alternative treatment methods
- ▶ State of the art is critical in assessing the benefit-risk ratio of the device, and you need to address it in your clinical evaluation report (CER).
- ▶ If your device has been on the market for decades and there are competitive devices that are technically superior and present lower risk than yours, they reduce the benefit of your device and increase the risk side of your benefit-risk equation.

# Paradigm shift in go-to-market strategy





# FDA's strategy to be first

## ► Received by a customer:

- “I am contacting submitters of Q-Subs ... to measure CDRH's performance in meeting one of its 2018-2020 strategic priorities:

“By December 31, 2020, more than 50% of manufacturers of novel technologies for the U.S. market intend to bring their devices to the U.S. first or in parallel with other major markets.”



# FDA's strategy to be first

## ► Received by a customer:

- “I am contacting submitters of Q-Subs ... to measure CDRH's performance in meeting one of its 2018-2020 strategic priorities:

“By December 31, 2020, more than 50% of manufacturers of novel technologies for the U.S. market intend to **bring their devices to the U.S. first or in parallel** with other major markets.”

### Result

**62% of novel technology manufacturers intend to bring their devices to the United States**

Based on an 87% response rate from companies contacted about novel technology devices between 2018 and 2020, 62% of companies intended to bring their devices to the United States first or in parallel with other major markets by December 31, 2020.

Source: [CDRH 2018-2020 Strategic Priorities Accomplishments](#)



# Free download

- Regulatory Pathway Assessment (RPA)
- Business Market Assessment (BMA)



[leanraqa.com/free-guides](https://leanraqa.com/free-guides)

## Questions?





**Your regulatory strategy**



**Your regulatory submissions**



**Your quality systems  
and compliance**



**Your audit management**



**Your due diligence**



**Your technical support**



**Your grief counseling**



lean **RAQA** ✓

**Your Michelle Lott, RAC**

[michelle@leanRAQA.com](mailto:michelle@leanRAQA.com)

520.275.9838

companies

large

mid-sized

small

startup

industries

medical device

biologics

pharma

cosmetics

food

dietary supplements